

# Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

CERTIFICATE OF STERILITY

Sterility Lot Number	Product Release Date:	Sterility Cert Lot
CS20200330-US	12-Dec-21	2100156B

Dose Info	Min Specified	Max Specified	Min Actual	Max Actual
(KgY)	35	50	38	47.8

Manufacturing Lot Number(s):	M2100094B
------------------------------	-----------

ITEM NUMBER(S)	SIZE(S)
CTL417CS	MD, LG, XL, 2X
CTL903CS	XL
CTL428CS	MD, LG, XL, 2X
CTL713CS	U

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of  $10^{-6}$ . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at [www.lakeland.com](http://www.lakeland.com) or contact Lakeland customer service at [info@lakeland.com](mailto:info@lakeland.com).

**M. LEE MORGAN, Ph.D., Director of Global Quality**  
Quality and Compliance

