

Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

CERTIFICATE OF STERILITY

Sterility Lot Number	Product Release Date:	Sterility Cert Lot
CS20220212-US	May 18, 2022	CR20220318

Dose Info	Min Specified	Max Specified	MIN Actual	MAX Actual
(kGy)	30.0	50.0	32	42.0

Manufacturing Lot Number(s):	M2200542 M2200543 M2200544 M2200545 M2200546 M2200547 M2200548 M2200549 M2200550 M2200551 M2200552 M2200553 M2200554 M2200555 M2200556 M2200557 M2200558 M2200559 M2200560 M2200561 M2200562 M2200563
------------------------------	---

ITEM NUMBER(S)	SIZE(S)
CTL191CS	MD, LG, XL, 2X, 3X
CTL428CS	MD, LG, 2X, 3X, 4X, 5X
CTL414CS	LG, XL, 2X
CTL417CS	MD, 2X, 3X, 5X
CTL903CSP	2X
CTL713CS	
CTL850CSP	18

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10⁻⁶. Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at staging.lakeland.com or contact Lakeland customer service at info@lakeland.com.

M. LEE MORGAN, Ph.D., Director of Global Quality
Quality and Compliance



Industrial Protective Clothing
Critical Environment
Fire Turnout Gear
Hi-Vis Safety Apparel



info@lakeland.com
lakeland.com
800-645-9291