

# Quality Certificate



## Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

**CERTIFICATE OF STERILITY**

Sterility Lot Number	Product Release Date:	Sterility Cert Lot
CS20220228-US	July 20, 2022	2201253-1(-2)

Dose Info	Min Specified	Max Specified	MIN Actual	MAX Actual
(kGy)	30.0	50.0	30	43.7

Manufacturing Lot Number(s):	M2200805 M2200806 M2200807 M2200808 M2200809 M2200810 M2200811 M2200812 M2200813 M2200814 M2200815 M2200816 M2200817 M2200818
------------------------------	---

ITEM NUMBER(S)	SIZE(S)
CTL191CS	MD, LG, XL, 2X, 3X, 4X
CTL414CS	LG, XL, 2X
CTL417CS	MD, LG, 2X, 3X, 5X

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of  $10^{-6}$ . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at [staging.lakeland.com](http://staging.lakeland.com) or contact Lakeland customer service at [info@lakeland.com](mailto:info@lakeland.com).

**M. LEE MORGAN, Ph.D., Director of Global Quality**  
Quality and Compliance



Industrial Protective Clothing  
Critical Environment  
Fire Turnout Gear  
Hi-Vis Safety Apparel



[info@lakeland.com](mailto:info@lakeland.com)  
[lakeland.com](http://lakeland.com)  
800-645-9291