

Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

Sterility Lot Number	Product Release Date:	Sterility Cert No.
CS20220303-US (2)	12th September 2022	2201614-2

Dose Info	Min Specified	Max Specified	Min Actual	Max Actual
(KgY)	30	50	30	45.7

Process Code	M2201073/1074/1075/1076
--------------	-------------------------

ITEM NUMBER(S)	SIZE(S)
CTL713CS	ONE SIZE
CTL850CSP-18	ONE SIZE
CTL903CSP	SMMS, 2X

Lakeland Industries Inc, doing business as "Lakeland Fire and Safety", certifies that the product/item number(s) and lot number(s) identified above have been processed and released in accordance with all relevant specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1 Sterilization of Health Care Products – Requirements for Validation and Routine Control – to provide a Sterility Assurance level of 10^{-6} . Do not use if package integrity has been compromised.

Product has a shelf-life of 5 years and will remain sterile for that period from the production date as long as the integrity of the packaging remains uncompromised.

Certificates of Sterility are available online www.lakeland.com or contact Customer Services at info@lakeland.com

D. Rumble, BSc, PCQI

Global Compliance and Product Manager

