

Quality Certificate



CERTIFICATE OF STERILITY

Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

Sterility Lot Number	Product Release Date:	Sterility Cert Lot
CS20220815-EU-2	December 13, 2022	2202116

Dose Info	Min Specified	Max Specified	MIN Actual	MAX Actual
(kGy)	27.5	42.0	27.5	38.1

Manufacturing Lot Number(s):	M2201428 M2201429 M2201430 M2201431 M2201432 M2201433 M2201434 M2201435 S2200306
------------------------------	---

ITEM NUMBER(S)	SIZE(S)
CTL428CS	SM, MD, LG, XL, 2X
CTL903CSP	SMMD, LGXL
CTL850CSP	18

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at staging.lakeland.com or contact Lakeland customer service at info@lakeland.com.

M. LEE MORGAN, Ph.D., Director of Global Quality
Quality and Compliance



Industrial Protective Clothing
Critical Environment
Fire Turnout Gear
Hi-Vis Safety Apparel



info@lakeland.com
lakeland.com
800-645-9291