

CERTIFICATE OF STERILITY

Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

| Sterility Lot Number | | Product Release Date: | | Sterility Cert Lot | |
|------------------------------|---------------|-----------------------|------------|--------------------|--|
| CS20221212-US | | March 23, 2023 | | 2300454-1 | |
| Dose Info | Min Specified | Max Specified | MIN Actual | MAX Actual | |
| (kGy) | 30.0 | 50.0 | 32.2 | 42.1 | |
| Manufacturing Lot Number(s): | M2300285 | | | | |
| ITEM NUMBER(S) | SIZE(S) | | | | |
| CTL428CS | XL | | | | |

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in **ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization** to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at staging.lakeland.com or contact Lakeland customer service at info@lakeland.com.

M. LEE MORGAN, Ph.D., Director of Global Quality
Quality and Compliance



Industrial Protective Clothing
Critical Environment
Fire Turnout Gear
Hi-Vis Safety Apparel



info@lakeland.com
lakeland.com
800-645-9291