

Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

CERTIFICATE OF STERILITY

Sterility Lot Number	Product Release Date:	Sterility Cert Lot
CS20230421-US	August 3, 2023	2301283-1, 2, 3, 4

Dose Info	Min Specified	Max Specified	MIN Actual	MAX Actual
(kGy)	27.5	42.0	30.8	40.4

Manufacturing Lot Number(s):	G2303174/3175
-------------------------------------	---------------

ITEM NUMBER(S)	SIZE(S)
SBC417CS	SM, MD, LG, XL, 2X, 3X, 4X, 5X
SBC414CS	3X, 4X, 5X
SBC191CS	SM, MD, LG, XL, 2X, 3X, 4X, 5X
SBC903CSP	SMMMD, LGXL, 2X
SBC713CS	
SBC850CSP	18
SBC428CS	SM, MD, LG, XL, 2X, 3X, 4X, 5X

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at staging.lakeland.com or contact Lakeland customer service at info@lakeland.com.

M. LEE MORGAN, Ph.D., Director of Global Quality
Quality and Compliance



Industrial Protective Clothing
Critical Environment
Fire Turnout Gear
Hi-Vis Safety Apparel



info@lakeland.com
lakeland.com
800-645-9291