

Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

CERTIFICATE OF STERILITY

Sterility Lot Number	Product Release Date:	Sterility Cert Lot
CS20230920-US	December 25, 2023	2400012

Dose Info	Min Specified	Max Specified	MIN Actual	MAX Actual
(kGy)	27.5	42.0	27.5	38.8

Manufacturing Lot Number(s):	M2301517/1518/1519/1520/1521/1522/1523/1524/1525/1526/1527/1528
------------------------------	---

ITEM NUMBER(S)	SIZE(S)
CTL191CS	LG
CTL414CS	MD-XL
CTL417CS	MD-LG-XL-2X
CTL428CS	LG-2X-5X
CTL903CSP	SMMD-LGXL

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in **ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization** to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at staging.lakeland.com or contact Lakeland customer service at info@lakeland.com.

M. LEE MORGAN, Ph.D., Director of Global Quality
Quality and Compliance



Industrial Protective Clothing
Critical Environment
Fire Turnout Gear
Hi-Vis Safety Apparel



info@lakeland.com
lakeland.com
800-645-9291