

# Quality Certificate



**CERTIFICATE OF STERILITY**

## Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

| Sterility Lot Number | Product Release Date: | Sterility Cert Lot |
|----------------------|-----------------------|--------------------|
| CS20231111-US        | January 26, 2024      | 2400194            |

| Dose Info | Min Specified | Max Specified | MIN Actual | MAX Actual |
|-----------|---------------|---------------|------------|------------|
| (kGy)     | 27.5          | 42.0          | 27.5       | 38.5       |

|                                     |  |
|-------------------------------------|--|
| <b>Manufacturing Lot Number(s):</b> | M2400098/0099/0100/0101/0102/0103/0104/0105/0106 |
|-------------------------------------|--|

| ITEM NUMBER(S) | SIZE(S)  |
|----------------|----------|
| CTL191CS       | LG       |
| CTL414CS       | MD-XL    |
| CTL417CS       | MD-LG-2X |
| CTL428CS       | 2X-4X-5X |

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of  $10^{-6}$ . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at [staging.lakeland.com](http://staging.lakeland.com) or contact Lakeland customer service at [info@lakeland.com](mailto:info@lakeland.com).

**M. LEE MORGAN, Ph.D., Director of Global Quality**  
Quality and Compliance



Industrial Protective Clothing  
Critical Environment  
Fire Turnout Gear  
Hi-Vis Safety Apparel



[info@lakeland.com](mailto:info@lakeland.com)  
[lakeland.com](http://lakeland.com)  
800-645-9291