

Quality Certificate



CERTIFICATE OF STERILITY

Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

Sterility Lot Number	Product Release Date:	Sterility Cert Lot
CS20231219-US	March 1, 2024	2400388-1

Dose Info	Min Specified	Max Specified	MIN Actual	MAX Actual
(kGy)	27.5	42.0	27.7	38.7

Manufacturing Lot Number(s):	M2400261/0262/0263/0264/0265/0266/0267
-------------------------------------	--

ITEM NUMBER(S)	SIZE(S)
CTL191CS	LG-2X
CTL414CS	LG-5X
CTL417CS	2X
CTL428CS	MD-4X

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at staging.lakeland.com or contact Lakeland customer service at info@lakeland.com.

M. LEE MORGAN, Ph.D., Director of Global Quality
Quality and Compliance



Industrial Protective Clothing
Critical Environment
Fire Turnout Gear
Hi-Vis Safety Apparel



info@lakeland.com
lakeland.com
800-645-9291