

**CERTIFICATE OF STERILITY**

# Quality Certificate



**Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel**

Sterility Lot Number		Product Release Date:		Sterility Cert Lot	
CS20240111-AP		January 26, 2024		2400195	
Dose Info		Min Specified	Max Specified	MIN Actual	MAX Actual
(kGy)		27.5	42.0	27.8	36.5
Manufacturing Lot Number(s):		M2400110			
ITEM NUMBER(S)		SIZE(S)			
CTL414CSD		XL			

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in **ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization** to provide a Sterility Assurance Level of  $10^{-6}$ . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at [staging.lakeland.com](http://staging.lakeland.com) or contact Lakeland customer service at [info@lakeland.com](mailto:info@lakeland.com).

**M. LEE MORGAN, Ph.D., Director of Global Quality**  
Quality and Compliance



Industrial Protective Clothing  
Critical Environment  
Fire Turnout Gear  
Hi-Vis Safety Apparel



[info@lakeland.com](mailto:info@lakeland.com)  
[lakeland.com](http://lakeland.com)  
800-645-9291