

Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

CERTIFICATE OF STERILITY

Sterility Lot Number	Product Release Date:	Sterility Cert Lot
CS20240221-US	05-Jun-24	2401132-2

Dose Info	Min Specified	Max Specified	Min Actual	Max Actual
(KgY)	27.5	42	27.6	41.2

Manufacturing Lot Number(s):	M2400979/0980/0981/0982/0983/0984/0985/0986/0987/0989/0990/0991/0992/0993/0994/0995/0996/0997/0998/0999/1000/100
-------------------------------------	--

ITEM NUMBER(S)	SIZE(S)
CTL417CS	MD, XL
CTL428CS	MD, LG, XL, 2X, 3X, 4X, 5X
CTL713CS	U
CTL903CSP	SMMD, LGXL

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.

D. Rumble, BSc, PCQI

Global Compliance and Product Manager

