

Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

CERTIFICATE OF STERILITY

Sterility Lot Number	Product Release Date:	Sterility Cert No.
CS20240932-US	20th October 2024	2402229-2

Dose Info	Min Specified	Max Specified	Min Actual	Max Actual
(KgY)	27.5	42	28.4	37.2

Process Code	M2401929/1930/1931/1932/1933/1934
--------------	-----------------------------------

ITEM NUMBER(S)	SIZE(S)
SBC428CS	XL, 2X, 3X, 4X, 5X
SBC903CSP	LGXL

Lakeland Industries Inc, doing business as "Lakeland Fire and Safety", certifies that the product/item number(s) and lot number(s) identified above have been processed and released in accordance with all relevant specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1 Sterilization of Health Care Products – Requirements for Validation and Routine Control – to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility are available online www.lakeland.com or contact Customer Services at info@lakeland.com

D. Rumble, BSc, PCQI

Global Compliance and Product Manager

