

Gamma Irradiation Certificate for CleanMax® Cleanroom Apparel

Certificate of Sterility

| Sterility Lot Number | Product Release Date | Sterility Certificate Number |
|----------------------|----------------------|------------------------------|
| CS20260116-US | 6th April 2026 | 2600806-2 |

| Dose Info | Min Specified | Max Specified | Min Actual | Max Actual |
|-----------|---------------|---------------|------------|------------|
| kGy | 27.5 | 42 | 27.6 | 40 |

| Process Code |
|-----------------------------------|
| M2600592/0593/0594/0595/0596/0597 |

| Item Number(s) | Size(s) |
|----------------|----------|
| CTL414CS | XL |
| CTL417CS | MD |
| CTL428CS | MD, LG |
| CTL713CS | ONE SIZE |
| CTL850CSP-18 | ONE SIZE |

Lakeland Industries Inc, doing business as "Lakeland Fire and Safety", certifies that the product/item number(s) and lot number(s) identified above have been processed and released in accordance with all relevant specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137 - 1 Sterilization of Health Care Products – Requirements for Validation and Routine Control – to provide a Sterility Assurance level of 10⁻⁶. Do not use if package integrity has been compromised.

Product has a shelf-life of 5 years and will remain sterile for that period from the production date as long as the integrity of the packaging remains uncompromised.

Certificates of Sterility are available online www.lakeland.com or contact Customer Services at info@lakeland.com



D. Rumble, BSc, PCQI

Global Compliance and Product Manager

