

Gamma Irradiation Certificate for CleanMax® Cleanroom Apparel

Certificate of Sterility

Sterility Lot Number	Product Release Date	Sterility Certificate Number
CS20260122-AP	23rd February 2026	2600433

Dose Info	Min Specified	Max Specified	Min Actual	Max Actual
kGy	27.5	42	27.5	39.2

Process Code	M2600344
--------------	----------

Item Number(s)	Size(s)
CTL428CS	MD

Lakeland Industries Inc, doing business as "Lakeland Fire and Safety", certifies that the product/item number(s) and lot number(s) identified above have been processed and released in accordance with all relevant specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137 - 1 Sterilization of Health Care Products – Requirements for Validation and Routine Control – to provide a Sterility Assurance level of 10^{-6} . Do not use if package integrity has been compromised.

Product has a shelf-life of 5 years and will remain sterile for that period from the production date as long as the integrity of the packaging remains uncompromised.

Certificates of Sterility are available online www.lakeland.com or contact Customer Services at info@lakeland.com



D. Rumble, BSc, PCQI

Global Compliance and Product Manager

